

**510(k) Summary  
for the Tiger Cannulated Screw System**

In accordance with 21 CFR 807.92 of the Federal Code of Regulations  
the following 510(k) summary is submitted for the Tiger Cannulated Screw System.

Date Prepared: May 27, 2008

- AUG - 4 2008**
1. **Submitter:**  
Trilliant Surgical Ltd  
1630 W.13th St  
Houston, TX 77008
  - Contact Person:**  
J.D. Webb  
The OrthoMedix Group, Inc.  
1001 Oakwood Blvd  
Round Rock, TX 78681  
Telephone: 512-388-0199
  2. **Trade name:** Tiger Cannulated Screw System  
**Common Name:** Bone screw  
**Classification Name:** Screw, Fixation, Bone  
Class II per 21 CFR section 888.3040  
HWC
  3. **Predicate or legally marketed devices which are substantially equivalent:**  
The Tiger Cannulated Screw System is substantially equivalent to similar previously cleared cannulated screws.
  4. **Description of the device:**  
The Tiger Cannulated Screw System is comprised of screws used for bone fixation of the hand and foot following trauma or osteotomy. The system features 2.0mm diameter, 2.4mm diameter, 3.0mm diameter and 4.0mm diameter cannulated screws. System instruments include 2.0mm/2.4mm drill bit, countersink, driver, 3.0mm/4.0mm drill bit, countersink, driver, screw driver handle, depth gauge, screw remover and K-wires to facilitate the placement of screws.  
  
**Materials:**  
The screws will be manufactured from titanium alloy (Ti-6Al-4V) per ASTM F136.  
  
**Function:**  
The Tiger Cannulated Screw Fixation System is comprised of screws used for bone fixation of the hand and foot following trauma or osteotomy.
  5. **Intended Use:**  
The Tiger Cannulated Screw Fixation System implants (screws) are intended for fixation of fractures, non-unions, arthrodeses and osteotomies of the small bones in the hand and foot.
  6. **Comparison of the technological characteristics of the device to predicate and legally marketed devices:**  
Due to the similarity of materials and design to both pre-enactment and post-enactment devices, Trilliant Surgical believes that the Tiger Cannulated Screw Fixation System does not raise any new safety or effectiveness issues and does not require any nonclinical testing.



Food and Drug Administration  
9200 Corporate Boulevard  
Rockville MD 20850

Trillian Surgical Ltd.  
% The OrthoMedix Group, Inc.  
Mr. J.D. Webb  
1001 Oakwood Boulevard  
Round Rock, Texas 78681

AUG - 4 2008

Re: K081510

Trade/Device Name: Tiger Cannulated Screw Fixation System  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth or threaded metallic bone fixation fastener  
Regulatory Class: II  
Product Code: HWC  
Dated: May 27, 2008  
Received: May 29, 2008

Dear Mr. Webb:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Center for Devices and Radiological Health's (CDRH's) Office of Compliance at (240) 276-0120. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding postmarket surveillance, please contact CDRH's Office of Surveillance and Biometric's (OSB's) Division of Postmarket Surveillance at (240) 276-3474. For questions regarding the reporting of device adverse events (Medical Device Reporting (MDR)), please contact the Division of Surveillance Systems at (240) 276-3464. You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at toll-free number (800) 638-2041 or (240) 276-3150 or the Internet address <http://www.fda.gov/cdrh/industry/support/index.html>.

Sincerely yours,



Mark N. Melkerson  
Director  
Division of General, Restorative  
and Neurological Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known): K081510 (pg 1/1)

Device Name: Tiger Cannulated Screw Fixation System

Indications for Use:

The Tiger Cannulated Screw Fixation System implants (screws) are intended for fixation of fractures, non-unions, arthrodeses and osteotomies of the small bones in the hand and foot.

Prescription Use X  
(Part 21 CFR 801 Subpart D)

AND/OR

Over-The-Counter Use \_\_\_\_\_  
(21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER PAGE IF NEEDED)

---

Concurrence of CDRH, Office of Device Evaluation (ODE)



**(Division Sign-Off)**  
**Division of General, Restorative,  
and Neurological Devices**

510(k) Number K081510



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

Trilliant Surgical, LTD  
% J.D. Webb  
President  
The OrthoMedix Group  
1001 Oakwood Boulevard  
Round Rock, Texas 78681

January 14, 2016

Re: K153338

Trade/Device Name: Tiger Cannulated Screw System  
Regulation Number: 21 CFR 888.3040  
Regulation Name: Smooth or threaded metallic bone fixation fastener  
Regulatory Class: Class II  
Product Code: HWC  
Dated: November 13, 2015  
Received: November 19, 2015

Dear Mr. Webb:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Parts 801); medical device reporting (reporting of medical device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set

forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to

<http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Industry and Consumer Education at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address

<http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,

**Mark N. Melkerson -S**

Mark N. Melkerson  
Director  
Division of Orthopedic Devices  
Office of Device Evaluation  
Center for Devices and  
Radiological Health

Enclosure

## Indications for Use

510(k) Number (if known)

K153338

Device Name

Tiger Cannulated Screw System

Indications for Use (Describe)

The Tiger Cannulated Screw System is indicated for fixation of fractures, non-unions, arthrodeses, and osteotomies of bones appropriate for the size of the device.

Type of Use (Select one or both, as applicable)

Prescription Use (Part 21 CFR 801 Subpart D)

Over-The-Counter Use (21 CFR 801 Subpart C)

**PLEASE DO NOT WRITE BELOW THIS LINE – CONTINUE ON A SEPARATE PAGE IF NEEDED.**

**FOR FDA USE ONLY**

Concurrence of Center for Devices and Radiological Health (CDRH) (Signature)

**510(k) Summary:**

In accordance with 21 CFR 807.92 of the Federal Code of Regulations

<b>Date Prepared</b>	November 13, 2015
<b>Submitted By</b>	Trilliant Surgical, LTD 6721 Portwest Dr, Suite 160 Houston, TX 77024 713-388-6055 Tele jolson@trilliantsurgical.com
<b>Contact</b>	J.D. Webb 1001 Oakwood Blvd Round Rock, TX 78681 512-388-0199 Tele e-mail: jdwebb@orthomedix.net
<b>Trade Name</b>	Tiger Cannulated Screw System
<b>Common Name</b>	Bone screws
<b>Classification Name</b>	Smooth & threaded metallic bone fixation fasteners
<b>Class</b>	II
<b>Product Code</b>	HWC
<b>CFR Section</b>	21 CFR section 888.3040
<b>Device Panel</b>	Orthopedic
<b>Primary Predicate Device</b>	Tiger Cannulated Screw System - K081510 / K112737
<b>Secondary Predicate Devices</b>	Synthes Cannulated Screws - K962011 / K963172 / K963192 / K021556 Osteomed ExtremiFix Cannulated Screw System - K063298 / K151021
<b>Device Description</b>	<p>The Tiger Cannulated Screw System is a headed and headless, cannulated, self-drilling, self-tapping screw system for the management of bone orthopedic osteotomies and trauma. The system consists of Ø2.0-Ø7.0 mm screws and washers, and the necessary instruments to facilitate the placement of these implants.</p> <p>The purpose of this submission is to add Ø5.5 mm and Ø7.0 mm headed cannulated screws to the Tiger Cannulated Screw System and to modify the indications for use statement for previously cleared devices (K081510 / K112737).</p>
<b>Materials</b>	Titanium alloy, Ti-6Al-4V (ASTM F136)
<b>Substantial Equivalence Claimed to Predicate Devices</b>	The Ø5.5 mm and Ø7.0 mm Tiger Cannulated Screws and Washers are substantially equivalent to the predicate devices in terms of intended use, design, materials used, mechanical safety and performances.



**510(k) Summary:**

In accordance with 21 CFR 807.92 of the Federal Code of Regulations

<b>Indications for Use</b>	The Tiger Cannulated Screw System is indicated for fixation of fractures, non-unions, arthrodeses, and osteotomies of bones appropriate for the size of the device.
<b>Non-clinical Test Summary</b>	<p>The following analyses were conducted:</p> <p><u>ASTM F543-07</u></p> <ul style="list-style-type: none"> <li>• Torque to Failure</li> <li>• Axial Pullout Strength</li> <li>• Axial Pull Through Strength</li> <li>• Driving/Insertion and Removal Torque</li> </ul> <p>The results of these evaluations indicate that the Tiger Cannulated Screws are equivalent to predicate devices.</p>
<b>Clinical Test Summary</b>	No clinical studies were performed
<b>Conclusions: Non-clinical and Clinical</b>	Trilliant Surgical considers the Ø5.5 mm and Ø7.0 mm Tiger Cannulated Screws and Washers to be equivalent to the predicate devices listed above. This conclusion is based upon the devices' similarities in principles of operation, technology, materials and indications for use